


Department of Health Central Luzon Center for Health Development

MARIVELES MENTAL WELLNESS AND GENERAL HOSPITAL





ct No: 09688545320 Email: procurement@mmwgh.gov.ph; procurement.mmwgh@gmail.com; Website: mmwgh

REQUEST FOR QUOTATION

Date:	September 5, 2023
PR No.:	2023-06-0396 (NUR)

						PR No.:	2023-06-0396 (NUR
ı. Business Nar	no:						
Address:	ile.						
Contact No./ E	-mail:						
IN:							
1							
Sir/ Madam:							
The Mariveles	s to quot			MMWGH) through its Bids and Awards Committee invite scribed herein subject to the Terms and Conditions and	-	-	• •
Required Doc	uments/	Information t	o be submitted	as attachments to the quotation/ Proposal			
PhilGEPS	Registrat	ion Number/ C	ertificate	 Tax Clearance 			
 Mayor's/ B 	usiness F	Permit		 Sample Invoice/ O.R./ C.R. 			
 DTI/ SEC 0 	Certificate)		 Omnibus Sworn Statement (for A 	ABCs above 50K; for	r winning bidder or	nly)
BIR Form 2	2303			 Income/ Business tax Return (fo 	r ABCs above 500K))	
) .							
his pro-forma	quotatio	n maybe subm	itted through reg	istered or electronic mail to the BAC Secretariat at above	e address or email to	procurement@mi	mwgh.gov.ph
r procuremen	t.mmwgh	@gmail.com c	on or before	11-Sep-2023	, 5PM subject t	o the following Te	ms and Conditions:
1 All entries	shall be t	vped or written	in a clear legible	manner			
			-	tive for THIRTY (30) calendar days upon issuance of thi	is document. Alterna	tive bids shall be r	eiected.
			-	eso shall include all applicable government taxes subjec			•
			ervices) deduction	**	,		,
4 MMWGH r	nay requi	re you to subn	nit additional doc	uments that will prove your legal, financial and technical	capability to underta	ke this contract	
5 Salient pro Penalties s			RR (RIRR) of Re	public Act (RA) 9184: Section 68 - Liquidated Damages	and Section 69 - Imp	position of Adminis	trative
6 MMWGH	reserves	the right to reje	ect any and all qu	otations, declare a failure, or not award the contract pur-	suant to Section 41	of the same RIRR	
7 In case of	tie quotat	ions, the BAC	shall adopt and e	employ "toss coin" as the tie-breaking method to finally de	etermine the single v	vinning provider in	accordance
with GPPB							
				conditions will be governed by the submitted signed Red			
-		-	•	I) working days upon receipt of Purchase Order (PO) or ompletion of delivery	within the working a	ays indicated in the	e PO.
-		iteen (10) won	ang days upon o	impletion of delivery			
Very Truly							
VINCENT	SGD.	AD A	_				
Chief Adm							
• Offici / tulli	iiiiottative	Olliooi					
). 		l	T		I	(to be filled up by th	e supplier)
Item No.	QTY	Unit	ABC	Technical Specfications/ Description	Brand/ Model	Unit Price	Total Amount
1	1	unit	55,000.00	Fetal Doppler			
				Ultrasound Frequency: 2MHZ			
				Ultrasound Intensity: <10mw/cm ²			
			ļ	Display: 45mm x 25mm LCD			
				FHZ Measuring Range: 50 - 240 bpm			
		ĺ		FHZ Resolution: 1 bpm			

item No.	QIT	Unit	ABC	Technical Specifications/ Description	Brand/ Model	Unit Price	Total Amount
1	1	unit	55,000.00	Fetal Doppler			
				Ultrasound Frequency: 2MHZ			
				Ultrasound Intensity: <10mw/cm ²			
				Display: 45mm x 25mm LCD			
				FHZ Measuring Range: 50 - 240 bpm			
				FHZ Resolution: 1 bpm			
				GHZ Accuracy: 1 bpm			
				Power Consumption: <1W			
				Main Body Dimension: 135mm x 95mm x 35mm			
				Power Supply Rechargeable Battery: ac220/110v			
				50/60HZ			
				Weight: 500g			
				See attached TOR for Details			
				Nothing Follows			
						Total Amount in	
						Figures	
						Total Amount in	
						Words	
		Total:	55,000.00				

Printed Name/Signature of Bidder

RFQ served by:

Canvasser



Central Luzon Center for Health Development

MARIVELES MENTAL WELLNESS AND GENERAL HOSPITAL

P. Mouroe Street, Poblacion, Mariveles, Bataan, Philippines, 2105

Crunkline: +63479354617; Office of the COH: +63476339006





TERMS OF REFERENCE

Name of Project

Supply, Delivery, Testing and Commissioning of Brand New **FETAL DOPPLER** (Public Bidding)

Technical Specification

- Ultrasound Frequency: 2MHZ

Ultrasound Intensity: <10mw/cm²

• Display: at least 45mm x 25mm LCD

- FHZ Measuring Range: 50 - 240 bpm

• FHZ Resolution: 1 bpm

- FHZ Accuracy: 1 bpm

- Power Supply Rechargeable Battery: 220VAC /60HZ

FUNCTION:

A test that uses soundwaves to check a baby's heartbeat. A handheld device to detect changes in movement that are translated as sound.





<u>MARIVELES MENTAL WELLNESS AND GENERAL HOSPITAL</u>

P. Mouroe Street, Poblacion, Mariveles, Bataan, Philippines, 2105





mmwgh.gov.ph


Trunkline: +63479354617: Office of the COH: +63476339006

mail@mmwgh.gov.ph

Documentary Requirements

- 1. Product brochure or technical data sheet(s) of the equipment showing the technical specification in English language.
- Valid and current Certificate of Compliance of manufacturer of the equipment with the latest version of ISO 13485: Quality Management System – Requirements for regulatory purposes in the name of the manufacturer. The Certificates must be issued by an independent Certifying Body/Agency.
- 3. Valid Marketing Authorization, Registration Approval or Free Sale Certificate for each equipment issued by the Health Authority in the country of origin.
- 4. Valid and current Certificate of Compliance and/or Test Report on the latest version of IEC 60601-1-1. The Certificate and/or Test Report must be issued by an independent Certifying Agency.
- Valid Certificate of Distribution (as first Tier Distributor) issued by the Manufacturer of each equipment authorizing the bidder to sell/distribute the offered equipment.
- 6. Proof (such as sales invoice) that the Brand of the equipment has been sold to other health facilities in the Philippines.
- 7. Notarized Certificate from the bidder:
 - a) That the brand of the equipment has been in the local and/or international market for at least ten (10) years.
 - That the equipment and its accessories are brand new, unused, not discontinued models and were not subjected to any product recall.
- 8. Bidder's valid and current License to Operate (LTO) as medical device distributor issued by the Philippine Food and Drug Administration. In case of expired LTO, the following must be submitted:
 - Copy of expired LTO
 - Application for renewal ii.
 - Official Receipt as proof of payment for the renewal of LTO iii.





MARIVELES MENTAL WELLNESS AND GENERAL HOSPITAL

P. Monroe Street, Poblacion, Marineles, Bataan, Philippines, 2105



mmwgh.gov.ph

Trunkline: +63479354617; Office of the COH: +63476339006

amail@mmwqh.qov.ph

Requirements if awarded the Contract

- Completion period: The delivery, testing and commissioning of the equipment and its accessories, including the training of end-user and maintenance staff must be completed within 45 calendar days upon receipt of the Notice to Proceed.
- 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.
- 3. Training: The supplier shall provide a training on proper use and maintenance of the equipment to the endusers and to the hospital maintenance staff.
- Warranty: Warranty certificate for two (2) years on parts and on services. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
- 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of equipment breakdown from the end-
- 6. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to warranty period.
- 7. Manuals: The supplier provide the end-user one (1) hard and one (1) soft copy of the following:
 - a. Service manual in English language
 - Operations manual in English language

Prepared by DIANA G! ALEJANDRO, RN Nurse III Head - Medical Ward

Approved b ZORAIDA F./AFABLE, MD Head, Medical Service BAC Chairperson

Attestation:

No item in the technical specifications and other requirements are reference to a specific brand of the equipment.

SGD.S MEYNARD ANTHONY V. BANZON, ECE

TWG-Healthcare Technology Management Section

